



Test Report issued under the responsibility of:



Prodigy Technology Consultant Co., Ltd.

**IEC / EN 60601-1**  
**Medical electrical equipment**  
**Part 1: General requirements for basic safety and essential performance**

Report Reference No. ....: M170518-06-A0

Date of issue .....: 2017-08-22

Total number of pages .....: 154

Testing Laboratory.....: Prodigy Technology Consultant Co., Ltd.

Address .....: No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan, R.O.C.

Applicant's name .....: AG Neovo

Address .....: 5F-1, No. 3-1, Park Street, Nangang District, Taipei, 11503, Taiwan

**Test specification:**

Standard.....: EN 60601-1:2006 + A11: 2011 + A1: 2013 + A12: 2014;  
IEC 60601-1: 2005 + CORR. 1:2006 + CORR. 2:2007 + AM1:2012 (or IEC 60601-1: 2012 reprint)

Test procedure .....: CB Scheme / Service in CE marking

Non-standard test method.....: None

Test Report Form No. ....: DTL-127-A2

Test Report Form Originator.....: Prodigy Technology Consultant Co., Ltd.

Master TRF.....: Dated 2014-07-07

Test item description .....: LCD Monitor

Trade Mark.....:



Manufacturer .....: TOP VICTORY ELECTRONICS (TAIWAN) CO LTD

Model/Type reference .....: MX-22x, MX-24x (where x can be A to Z (a-z), 0 to 9 or "+", "-", "\", "/" )

Ratings .....: Power Adapter: Adapter Technology Co., Ltd. / ATM065-P120 or ATM065T-P120  
Input: 100-240Vac, 50-60Hz, 1.6-0.7A;  
Output: 12Vdc, 5.0A  
LCD Monitor rating: 12Vdc, 2.5A