



Date : May 06, 2026

CE Statement of Conformity

This statement is to certify that the designated product below.

Product : Medical Panel PC
Trademark : AG Neovo
Model Number : TXA2401
Company Name : Associated Industries China, Inc.
Applicable Standards : EN 60601-1-2 2015/A1:2021, Group1, Class B
EN IEC 61000-3-2:2019/A2:2024
EN 61000-3-3:2013/A2:2021/AC:2022-01

One sample of the designated product has been tested and evaluated in our laboratory to find in compliance with the applicable standards above. The issued test report(s) show(s) it in detail.

Report Number : 2630445R-0E3012310006-A

TEST LABORATORY

A handwritten signature in black ink, appearing to read "Vincent Lin".

Vincent Lin / Director

The verification is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.